

Step by step guide for the grinding of medication with the Severo Professional Medication Grinder

Follow this recommended grinding protocol to avoid unwanted exposure to medication dust.
Suivez ce protocole de broyage recommandé pour éviter une exposition indésirable à de la poudre médicamenteuse.
Siga este protocolo de molido para evitar una exposición no deseada a polvo de medicamento.
Befolgen Sie dieses empfohlene Zerkleinerungsprotokv, um sich nicht ungewollt Medikamentenstaub auszusetzen.

To avoid the risk of exposure and cross contamination please follow the recommended protocol as shown below. Alternatively check the following tutorial online at www.severo.nl/demovideo.



1 Open the Severo by pushing the bottom part of the grey handle. Put the medication in the Severo Cup, **fill it no more than the first line** and place a Severo Rondel centered on top.



2 Place the metal grinding head on top, push it in the cup and place it all in the Severo.



3 Close the lid of the Severo. When the device is in standby-mode it immediately starts grinding. Otherwise, press the on/off button to start the grinding process. When the grinding cycle is completed, the on/off button will light up green.



4 After grinding, open the lid and take out the SeveroCup together with the grinding head. Check the cup for visual contaminations. Remove the grinding head while keeping the SeveroRondel in the cup.



5 Check the grinding head for visual contaminations. Gently remove the SeveroRondel, check for damages and tap off any remaining medication before disposal.



6 The grinded medication in the cup can be mixed with a liquid or with medication gel before administration. The grinding head is ready for re-use.

Requirements

• **SeveroCup** (article code: MED001)

The following additional items are recommended for use during medication grinding with the Severo:

• **SeveroRondel** (article code: MED003)

For keeping the grinding head clean. The use of rondels prevents cross contamination and saves cleaning time after each grinding cycle.

• **SeveroCup Lid** (article code: MED002)

For completely sealing off the cup, this way minimizing spills and contamination risks during storage and medication distribution.

Cleaning

Use an alcohol cloth to remove visible contamination on the grinding head. Visible contaminations on the Severo device can best be cleaned with a damp cloth and some detergent.

Tips

Always check the SeveroCup and SeveroRondel for damages after grinding. The disposables have been extensively tested with the hardest kinds of medication. However under extreme conditions damages can not fully be ruled out.

To maximise the lifespan of the battery, use the Severo disconnected from the recharger.

Always consult your physician or pharmacist before grinding any medication.

High Quality Medical Product

ISO 2001: 2015



LOT: Rb
C2017

Producer:

Samarit Medical AG

CH-8132 Egg / ZH

www.samarit.com

Tel. +41-44-918 10 11

e-mail: info@samarit.com